

Latex Particle Challenge Final Report

Test Article: Bioscarf 1 - Black Study Number: 973987-S01 Study Received Date: 29 Jun 2017 Nelson Laboratories, LLC, a Business Unit of Sterigenics International Testing Facility: 6280 S. Redwood Rd. Salt Lake City, UT 84123 U.S.A. Test Procedure(s): Standard Test Protocol (STP) Number: STP0005 Rev 05

Summary: This procedure was performed to evaluate the non-viable particle filtration efficiency (PFE) of the test article. Monodispersed polystyrene latex spheres (PSL) were nebulized, dried, and passed through the test article. The particles that passed through the test article were enumerated using a laser particle counter.

Three one-minute counts were performed, with the test article in the system, and the results averaged. Three one-minute control counts were performed, without a test article in the system, before and after each test article and the counts were averaged. Control counts were performed to determine the average number of particles delivered to the test article. The filtration efficiency was calculated using the average number of particles penetrating the test article compared to the average of the control values.

The procedure employed the basic particle filtration method described in ASTM F2299, with some exceptions; notably the procedure incorporated a non-neutralized challenge. In real use, particles carry a charge, thus this challenge represents a more natural state. The non-neutralized aerosol is also specified in the FDA guidance document on surgical face masks. All test method acceptance criteria were met. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.

Test Side: Either Side Area Tested: 91.5 cm² Particle Size: 0.1 µm Laboratory Conditions: 21°C, 33% relative humidity (RH) at 1050; 21°C, 33% RH at 1232 Average Filtration Efficiency: 99.75% Standard Deviation: 0.044





Study Completion Date

973987-S01 P.O. Box 571830 | Murray, UT 84157-1830 U.S.A. 6280 South Redwood Road | Salt Lake City, UT 84123-6600 U.S.A. www.nelsonlabs.com · Telephone 801 290 7500 · Fax 801 290 7998 · sales@nelsonlabs.com

FRT0005-0001 Rev 4 Page 1 of 2

These results relate only to the test article listed in this report. Reports may not be reproduced except in their entirety. Subject to NL terms and conditions at www neisonlabs com

Study Number 973987-S01 Latex Particle Challenge Final Report



Results:

Test Article Number	Average Test Article Counts	Average Control Counts	Filtration Efficiency (%)
1	35	11,436	99.69
2	26	11,075	99.77
3	28	11,083	99.75
4	20	10,970	99.82
5	28	11,074	99.74

P.O. Box 571830 | Murray, UT 84157-1830 U.S.A. + 6280 South Redwood Road | Salt Lake City, UT 84123-6600 U.S.A. www.nelsonlabs.com + Telephone 801 290 7500 + Fax 801 290 7998 + sales@nelsonlabs.com

lbv